

Ellipse™ DR

Dual-Chamber Implantable Cardioverter Defibrillator (ICD)

Product Highlights

- MRI-ready device to allow patients to safely undergo an MRI scan when used in combination with MR Conditional leads^{1,2}
- Improved shape with reduced volume and thickness
- Parylene coating for improved abrasion resistance
- DynamicTx™ Over-Current Detection Algorithm automatically changes shock configurations to ensure delivery of high voltage therapy when high current is detected
- Cold Can programmability provides an additional RV-SVC Shock Configuration to decouple the can from the shocking vector parameters in cases of lead problems
- ShockGuard™ technology with DecisionTx™ programming designed to reduce inappropriate therapy and minimise the need for programming adjustments at implant
- Low Frequency Attenuation filter designed to enhance sensing performance and may reduce the possibility of oversensing T-waves
- SenseAbility™ feature provides flexibility to fine-tune programming around T-wave oversensing without decreasing sensitivity
- DF4 connector designed to streamline defibrillation connections into a single terminal pin and reduce the number of set screws
- CorVue™ congestion monitoring feature monitors the intrathoracic impedance in multiple vectors for improved accuracy, and it provides the option for both patient and physician alerts
- SecureSense™ RV lead noise discrimination detects sustained and short bursts of lead noise that would otherwise go unnoticed or potentially lead to one or more inappropriate shocks
- Far Field MD™ morphology discrimination and Chamber Onset discrimination improve SVT and VT discrimination for reduced inappropriate therapies
- Antitachycardia pacing (ATP) while charging and prior to charging in the VF zone further extends the programming options for terminating tachyarrhythmias without a high-voltage shock
- ST monitoring capability provides unprecedented, continuous insight into significant ST shift events and associated ventricular arrhythmias through enhanced monitoring of iEGM and ST-segment as a diagnostic tool to help guide appropriate clinical action
- 36 J delivered energy safety shock option can provide a greater DFT safety margin
- DeFT Response™ technology offers the most noninvasive options for managing high DFTs
- QHR™† chemistry battery provides greater capacity for enhanced longevity and improved charge time performance compared to previous SVO batteries



Merlin@home™
Transmitter
Compatible



Ordering Information

DRAFT SPECIFICATIONS; CE MARK PENDING

Contents: Cardiac pulse generator

Model Number	Dimensions (H x W x T, mm)	Weight (g)	Volume (cc)	Connector Defibrillation	Connector Sense/Pace
CD2377-36C	69 x 51 x 12	66	31	DF1	IS-1
CD2377-36QC*	70 x 51 x 12	68	31	DF4	IS-1; DF4

*Indicates models that are MRI Conditional^{1,2}

Indications: The devices are intended to provide ventricular antitachycardia pacing and ventricular defibrillation for automated treatment of life-threatening ventricular arrhythmias.

Contraindications: Contraindications for use of the pulse generator system include ventricular tachyarrhythmias resulting from transient or correctable factors such as drug toxicity, electrolyte imbalance, or acute myocardial infarction.

Adverse Events: Implantation of the pulse generator system, like that of any other device, involves risks, some possibly life-threatening. These include but are not limited to the following: acute hemorrhage/bleeding, air emboli, arrhythmia acceleration, cardiac or venous perforation, cardiogenic shock, cyst formation, erosion, exacerbation of heart failure, extrusion, fibrotic tissue growth, fluid accumulation, hematoma formation, histotoxic reactions, infection, keloid formation, myocardial irritability, nerve damage, pneumothorax, thromboemboli, venous occlusion. Other possible adverse effects include mortality due to: component failure, device-programmer communication

failure, lead abrasion, lead dislodgment or poor lead placement, lead fracture, inability to defibrillate, inhibited therapy for a ventricular tachycardia, interruption of function due to electrical or magnetic interference, shunting of energy from defibrillation paddles, system failure due to ionising radiation. Other possible adverse effects include mortality due to inappropriate delivery of therapy caused by: multiple counting of cardiac events including T waves, P waves, or supplemental pacemaker stimuli. Among the psychological effects of device implantation are imagined pulsing, dependency, fear of inappropriate pulsing, and fear of losing pulse capability.

Refer to the User's Manual for detailed indications, contraindications, warnings, precautions and potential adverse events.

†QHR is a trademark of Greatbatch Medical

Ellipse™ DR

Dual-Chamber Implantable Cardioverter Defibrillator (ICD)

Product Specifications

PHYSICAL SPECIFICATIONS

Models	CD2377-36C	CD2377-360C
Telemetry	RF	RF
Delivered/Stored Energy (J)	36/39	36/39
Volume (cc)	31	31
Weight (g)	66	68
Size (mm)	69 x 51 x 12	70 x 51 x 12
Defibrillation Lead Connections	DF1	DF4
Sense/Pace Lead Connections	IS-1	IS-1; DF4
High-Voltage Can	Electrically active titanium can	Electrically active titanium can
Coating	Parylene	Parylene
MRI Conditional	No	Yes-MRI-ready

PARAMETER SETTINGS

AF Management

AF Suppression™ Pacing	On; Off
No. of Overdrive Pacing Cycles	15-40 in steps of 5
Maximum AF Suppression Rate	80-150 min ⁻¹

Sensing/Detection

SenseAbility™ Technology	Automatic Sensitivity Control adjustment for atrial and ventricular events
Low Frequency Attenuation	On; Off
Threshold Start	(Post-Sensed; Atrial) 50; 62.5; 75; 100%; (Post-Paced; Atrial) 0.2-3.0 mV; (Post-Sensed; Ventricular) 50; 62.5; 75; 100%; (Post-Paced; Ventricular) Auto; 0.2-3.0 mV
Decay Delay	(Post-Sense/Post-Pace; Atrial/Ventricular) 0-220
Ventricular Sense Refractory (ms)	125; 157
Detection Zones	3 zone programming - 1 zone, 2 zones or 3 zones (VT-1, VT-2, VF)
SVT Discriminators	AV Rate Branch; Arrhythmia Onset (Chamber Onset or Sudden Onset); Interval Stability; AV Association; Morphology Discrimination (Far Field MD or Original MD) with Manual (original MD only) or Automatic Template Update
Monitor Mode	Detection, discrimination and diagnostics, no therapy delivery (VT or VT-1 zone)
Discrimination modes	On; Passive; Off
SVT Threshold	150-240 min ⁻¹
SVT Timeout	0.25-5 min
Reconfirmation	Continuous sensing during charging
Lead Noise Discrimination	SecureSense™ RV lead noise discrimination (On; On with Timeout; Passive; Off)

Antitachycardia Pacing Therapy

ATP Configurations	Ramp; Burst; Scan; 1 or 2 schemes per VT zone
ATP in VF Zone	ATP While Charging; ATP Prior to Charging; Off
ATP Upper Rate Cutoff	150 - 300 min ⁻¹
Burst Cycle Length	Adaptive; Readaptive or Fixed
Min. Burst Cycle Length (ms)	150-400 in increments of 5
Number of Bursts	1-15
Number of Stimuli	2-20
Add Stimuli per Burst	On; Off
ATP Pulse Amplitude (V)	7.5 Independent from Bradycardia and Post-Therapy Pacing
ATP Pulse Width (ms)	1.0 or 1.5 Independently programmable from Bradycardia and Post-Therapy Pacing

High-Voltage Therapy

DynamicIX™ Algorithm	On; Off
DeFT Response™ Technology	Programmable pulse width for P1/P2 and tilt
High-Voltage Output Mode	Fixed Pulse Width; Fixed Tilt
Waveform	Biphasic; Monophasic
RV Polarity	Cathode (-); Anode (+)
Electrode Configuration	RV to Can; RV to SVC/Can; RV to SVC

Bradycardia Pacing

Permanent Modes	Off; DDD(R); DDI(R); VVI(R); AAI(R)
Temporary Modes	Off; DDD; DDI; VVI; AAI; AAT; DOO; VOO; AOO
Rate-Adaptive Sensor	On; Off; Passive
Programmable Rate and Delay Parameters	Base Rate (min ⁻¹); Rest Rate (min ⁻¹); Maximum Tracking Rate (min ⁻¹); Off; Maximum Sensor Rate (min ⁻¹); Paced AV Delay (ms); Sensed AV Delay (ms); Rate Responsive AV Delay (Atrial and RV) (ms); Hysteresis Rate (min ⁻¹); Rate Hysteresis with Search
Ventricular AutoCapture™	
Pacing System	On; Off
ACap™ Confirm	On; Monitor; Off
QuickOpt™ Timing Cycle Optimisation	Sensed/Paced AV delay

Auto Mode Switch (AMS)	Off; DDI(R); VVI(R)
Atrial Tachycardia	
Detection Rate (min ⁻¹)	110-300
AMS Base Rate (min ⁻¹)	40; 45;...135
Auto PMT Detection/Termination	Atrial Pace on PMT; Off; Passive
Rate Responsive PVARP/VREF	Off; Low; Medium; High
Ventricular Intrinsic Preference (VIP™)	Off; On (50-200)

Post-Therapy Pacing (Independently programmable from Bradycardia and ATP)

Post-Shock Pacing Mode	Off; AAI; VVI; DDI; DDD
Post-Shock Base Rate (min ⁻¹)	30-100 in increments of 5
Post-Shock Pacing Duration (min)	Off; 0.5; 1; 2.5; 5; 7.5; or 10

Device Testing/Induction Methods

DC Fiber™ Pulse Duration (sec)	0.5-5.0
Burst Fiber Cycle Length (ms)	20-100
Noninvasive Programmed Stimulation (NIPS)	2-25 stimuli with up to three extrastimuli

Patient Notifiers

Programmable Notifiers (On; Off)	Device at ERI; Charge Time Limit Reached; Possible HV Circuit Damage; Atrial Lead Impedance Out of Range; Ventricular Lead Impedance Out of Range; High-Voltage Lead Impedance Out of Range; AT/AF Burden; V Rate During AT/AF; % V pacing; CorVue Congestion Trigger; SecureSense – lead noise detected, non-sustained lead noise detected, ST Episodes (Type I only)
Device Parameter Reset	On
Entry into Backup VVI Mode	On
Vibration Duration (sec)	2; 4; 6; 8; 10; 12; 14; 16
Number of Vibrations per Notification	2
Number of Notifications	1-16
Time Between Notifications (hours)	10; 22

Electrograms and Diagnostics

Stored Electrograms	Up to 45 minutes including up to one minute programmable pre-trigger data per VT/VF diagnosis/detection electrograms; triggers include diagnosis; detection; therapy; atrial episode; PMT termination; PC shock delivery; noise reversion; magnet reversion; morphology template verification; lead noise detected, non-sustained lead noise detected, NSVT/NSVF
Therapy Summary	Diagram of therapies delivered
Episodes Summary	Directory listing of up to 60 episodes with access to more details including stored electrograms
Lifetime Diagnostics	History of bradycardia events and device-initiated charging
AT/AF Burden Trend	Trend data and counts
Ventricular HV Lead Impedance Trend	Multi-Vector Trend Data
Histograms	Event Histogram; AV Interval Histogram; Mode Switch Duration Histogram; Peak Filtered Rate Histogram; Atrial Heart Rate Histogram; Ventricular Heart Rate Histogram; AT/AF Burden; Exercise and Activity Trending; V Rates during AMS; DirectTrend™ reports up to 1 year
PMT Data	Information regarding PMT detections
Real-Time Measurements (RTM)	Pacing lead impedances; high-voltage lead impedances; and signal amplitudes
ST Monitoring	ST Histogram Data; Long-term ST Deviation Trend; ST Episode Log; ST Episode Details; 24-Hour ST & HR Trend; ST EGM Baseline & Snapshots prior to ST Episode, VT/VF, Interrogation (Snapshots and 24-hour trend at time of interrogation)
CorVue™ Congestion Monitoring	On; Off
CorVue Congestion Trigger	8-18 days

1. MRI Conditional Parameters: 1.5 Tesla, 2 W/Kg SAR

2. See MRI Procedure Information for approved MR Conditional Systems Device/Lead combinations and scan parameters

Customer Support: 46-8-474-4756

Brief Summary: Prior to using these devices, please review the Instructions for Use for a complete listing of indications, contraindications, warnings, precautions, potential adverse events and directions for use. Devices depicted may not be available in all countries. Check with your St. Jude Medical representative for product availability in your country.

Unless otherwise noted, ™ indicates that the name is a trademark of, or licensed to, St. Jude Medical or one of its subsidiaries. ST, JUDE MEDICAL and the nine-squares symbol are registered and unregistered trademarks and service marks of St. Jude Medical, Inc. and its related companies. ©2014 St. Jude Medical, Inc. All rights reserved.

SJM-ELP-1114-0002 | This document is for international use only.