# Quadra Assura<sup>™</sup>

## Cardiac Resynchronisation Therapy Defibrillator (CRT-D)

# Product Highlights

- The Quadra Assura CRT-D and Quartet™ quadripolar LV pacing lead feature four pacing electrodes and 10 pacing vectors to provide more options and greater control to minimise implant complications such as diaphragmatic stimulation and high pacing thresholds
- VectSelect Quartet<sup>™</sup> multivector testing feature offers a streamlined workflow to identify, test and program the patient's pacing vector
- DynamicTx<sup>™</sup> Over-Current Detection Algorithm automatically changes shock configurations to ensure delivery of high voltage therapy when high current is detected
- Parylene coating for improved abrasion resistance
- Cold Can programmability provides an additional RV-SVC Shock Configuration to decouple the can from the shocking vector parameters in cases of lead problems
- ShockGuard™ technology with DecisionTx™ programming designed to reduce inappropriate therapy and minimise the need for programming adjustments at implant
- SecureSense™ RV lead noise discrimination detects sustained and short bursts of lead noise that would otherwise go unnoticed or potentially lead to one or more inappropriate shocks
- Far Field MD™ morphology discrimination and Chamber Onset discrimination improve SVT and VT discrimination for reduced inappropriate therapies
- Antitachycardia pacing (ATP) while charging and prior to charging in the VF zone further extends the programming options for terminating tachyarrhythmias without a high-voltage shock
- Low Frequency Attenuation filter designed to enhance sensing performance and may reduce the possibility of oversensing T-waves
- SenseAbility™ feature provides flexibility to fine-tune programming around T-wave oversensing without decreasing sensitivity
- DF4 connector designed to streamline defibrillation connections into a single terminal pin and reduce the number of set screws
- Unique 40 J delivered energy safety shock option can provide a greater DFT safety margin
- DeFT Response<sup>™</sup> technology offers the most noninvasive options for managing high DFTs
- QHR<sup>™\*</sup> chemistry battery provides greater capacity for enhanced longevity and improved charge time performance compared to previous SVO batteries
- Vibratory patient notifier enables patients with hearing problems to be alerted to a low battery, lead-related complications and more
- CorVue<sup>™</sup> congestion monitoring feature monitors the intrathoracic impedance in multiple vectors for improved accuracy, and it provides the option for both patient and physician alerts
- QuickOpt™ timing cycle optimisation provides quick and effective optimisation at the push of a button

# Ordering Information

Contents: Cardiac pulse generator

Model Number	Dimensions (H x W x T, mm)	Weight (g)	Volume (cc)	Connector
CD3367-40C	83 x 41 x 14	83	40	DF1, IS4, IS-1
CD3367-40QC	75 x 41 x 14	80	38	DF4, IS4, IS-1

Indications: The devices are intended to provide ventricular antitachycardia pacing and ventricular defibrillation for automated treatment of life-threatening ventricular arrhythmias. Cardiac Resynchronisation Therapy Defibrillators (CRT-Ds) are also intended to resynchronise the right and left ventricles in patients with congestive heart failure.

Contraindications: Contraindications for use of the pulse generator system include ventricular tachyarrhythmias resulting from transient or correctable factors such as drug toxicity, electrolyte imbalance, or acute myocardial infarction.

Adverse Events: Implantation of the pulse generator system, like that of any other device, involves risks, some possibly life-threatening. These include but are not limited to the following: acute hemorrhage/bleeding, air emboli, arrhythmia acceleration, cardiac or venous perforation, cardiogenic shock, cyst formation, erosion, exacerbation of heart failure, extrusion, fibrotic tissue growth, fluid accumulation, hematoma formation, histotoxic reactions, infection, keloid formation, myocardial irritability, nerve damage, pneumothorax,

thromboemboli, venous occlusion. Other possible adverse effects include mortality due to: component failure, device-programmer communication failure, lead abrasion, lead dislodgment or poor lead placement, lead fracture, inability to defibrillate, inhibited therapy for a ventricular tachycardia, interruption of function due to electrical or magnetic interference, shunting of energy from defibrillation paddles, system failure due to ionising radiation. Other possible adverse effects include mortality due to inappropriate delivery of therapy caused by: multiple counting of cardiac events including T waves, P waves, or supplemental pacemaker stimuli. Among the psychological effects of device implantation are imagined pulsing, dependency, fear of inappropriate pulsing, and fear of losing pulse capability.

Refer to the User's Manual for detailed indications, contraindications, warnings, precautions and potential adverse events.

\*QHR is a trademark of Greatbatch Medical







# Quadra Assura<sup>™</sup>

## Cardiac Resynchronisation Therapy Defibrillator (CRT-D)

# **Product Specifications**

PHYSICAL SPECIFICATIONS				
Models	CD3367-40C	CD3367-40QC		
Telemetry	RF	RF		
Delivered/Stored Energy (J)	40/45	40/45		
Volume (cc)	40	38		
Weight (g)	83	80		
Size (mm)	83 x 41 x 14	75 x 41 x 14		
Defibrillation Lead Connections	DF1	DF4-LLHH		
LV Lead Connections	IS4-LLLL	IS4-LLLL		
Sense/Pace Lead Connections	IS-1	IS-1		
High-Voltage Can Coating	Electrically active titanium can Parylene	Electrically active titanium can Parylene		
	•	1 aryletie		
PARAMETER	SETTINGS			
Biventricular Pacing				
VectSelect Quartet™ LV  V. Triggering	Distal Tip 1 - Mid 2, Distal Tip 1 - Proximal 4, Distal Tip 1 - RV Coil; Pulse Configuration Mid 2 - Proximal 4; Mid 2 - RV Coil; Mid 3 - Mid 2; Mid 3 - PV Coil; Mid 3 - RV Coil; Proximal 4 - Mid 2; Proximal 4 - RV Coil On; Off			
QuickOpt™ Timing				
Cycle Optimisation		Sensed/paced AV delay, interventricular pace delay		
V-V Timing	Simultaneous*; RV First; LV First			
Interventricular Pace Delay (ms)	RV First 10-80 / LV First 15-80 in increments of 5			
Ventricular Sensing		RV only (not programmable)		
Ventricular Pacing Chamber		RV only; biventricular		
Negative AV Hysteresis/Search (ms)	Off; -10 to -120			
Shortest AV Delay (ms)	25-120			
AF Management				
AF Suppression™ Pacing No. of Overdrive Pacing Cycles Maximum AF Suppression Rate	On; Off 15-40 in steps of 5 80-150 min <sup>-1</sup>			
Sensing/Detection				
Sense <i>Ability</i> ™ Technology	Automatic Sensitivity Control adjustm and ventricular events	ent for atrial		
Low Frequency Attenuation	On; Off			
Sense Filter	(Post-Sensed; Atrial) 50; 62,5; 75; 100	)%: (Post-Paced: Atrial)		
	0,2-3,0 mV; Threshold Start (Post-Sensed; Ventricular)			
	50; 62,5; 75; 100%; (Post-Paced; Vent			
Decay Delay	(Post-Sensed/Post-Paced; Atrial/Vent			
Ventricular Sense Refractory (ms)	125; 157	11041417 0 220		
Detection Zones		3 zone programming - 1 zone, 2 zones or 3 zones (VT-1, VT-2, VF)		
SVT Discriminators	AV Rate Branch; Arrhythmia Onset (Chamber Onset or Sudden Onset);			
OVI Discriminators	Interval Stability; AV Association; Mor			
	(Far Field MD or Original MD) with Mar			
	or Automatic Template Update			
Monitor Mode	Detection, discrimination and diagnos (VT or VT-1 zone)	tics, no therapy delivery		
Discrimination modes				
	On, Paccive, Off			
	On; Passive; Off 150-240 min-1			
SVT Threshold	150-240 min <sup>-1</sup>			
SVT Threshold SVT Timeout	150-240 min <sup>-1</sup> 0; 25-5 min			
SVT Threshold	150-240 min <sup>-1</sup>	uation (On: On with Timeout: Pas-		

Antitachycardia Pacing Therapy ATP Configurations ATP in VF Zone Ramp; Burst; Scan; 1 or 2 schemes per VT zone ATP While Charging; ATP Prior to Charging; Off ATP Upper Rate Cutoff Burst Cycle Length 150-300 min<sup>-1</sup> Adaptive; Readaptive or Fixed Min. Burst Cycle Length (ms) 150-400 in increments of 5 1-15 with 2-20 Stimuli Number of Bursts/Stimuli Add Stimuli per Burst ATP Pulse Amplitude (V) 7,5 Independent from Bradycardia and Post-Therapy Pacing ATP Pulse Width (ms) 1,0 or 1,5 Independently programmable from Bradycardia and Post-Therapy Pacing

## High-Voltage Therapy

 $DynamicTx^{\scriptscriptstyle\mathsf{TM}}\ Algorithm$ On; Off DeFT Response™ Technology Programmable pulse width for P1/P2 and tilt High-Voltage Output Mode Fixed Pulse Width; Fixed Tilt Waveform Biphasic; Monophasic Cathode (-); Anode (+) Electrode Configuration RV to Can; RV to SVC/Can; RV to SVC

### Bradycardia Pacing

Permanent Modes Off; DDD(R); DDT(R); DDI(R); VVT(R); VVI(R); AAI(R)
Off; DDD; DDT; DDI; VVT; VVI; AAI; AAT; DOO; VOO; AOO Temporary Modes Rate-Adaptive Sensor Programmable Rate and

**Delay Parameters** Off; Base Rate (min-1); Rest Rate (min-1); Maximum Tracking Rate (min-1); Maximum Sensor Rate (min<sup>-1</sup>); Paced AV Delay (ms); Sensed AV Delay (ms); Rate Responsive AV Delay; Hysteresis Rate (min<sup>-1</sup>); Rate Hysteresis

BiVCap™ Confirm; LVCap™ Confirm;

RVCap™ Confirm ACap™ Confirm Setup; On; Monitor; Off Setup: On: Monitor: Off QuickOpt™ Timing Cycle Optimisation Interventricular Pace Delay Off; DDI(R); DDT(R); VVI(R); VVT(R)Auto Mode Switch (AMS) Atrial Tachycardia

Detection Rate (min-1) AMS Base Rate (min-1) 110-300 40; 45; ... 135 Auto PMT Detection/Termination Atrial Pace; Off; Passive Rate Responsive PVARP/VREF Off; Low; Medium; High Ventricular Intrinsic Preference (VIPT Off; On (50-200)

### Post-Therapy Pacing (independently programmable from Bradycardia and ATP)

Post-Shock Pacing Mode Off; AAI; VVI; DDI; or DDD Post-Shock Base Rate (min-1) 30-100 in increments of 5 Off; 0,5; 1; 2,5; 5; 7,5; or 10 Post-Shock Pacing Duration (min)

### **Device Testing/Induction Methods**

DC Fibber™ Pulse Duration (sec) 0.5 - 5.0Burst Fibber Cycle Length (ms)

Noninvasive Programmed Stimulation (NIPS) 2-25 stimuli with up to 3 extrastimuli

### **Patient Notifiers**

Programmable Notifiers (On; Off)

Device at ERI; Charge Time Limit Reached; Possible HV Circuit Damage; Atrial Lead Impedance Out of Range; RV Lead Impedance Out of Range; LV Lead Impedance Out of Range; High-Voltage Lead Impedance Out of Range; AT/AF Burden; V Rate During AT/AF; AT/AF Episode Duration; % V Pacing: CorVue™ Congestion Trigger, SecureSense — lead noise detected, non-sustained lead noise detected

Device Parameter Reset Entry into Backup VVI Mode

Vibration Duration (sec) 2; 4; 6; 8; 10; 12; 14; 16 Number of Vibrations per Notification Number of Notifications 1-16 Time Between Notifications (hours) 10; 22

## **Electrograms and Diagnostics**

Stored Electrograms

PMT Data

Up to 45 minutes; including up to 1 minute programmable pre-trigger data per VT/VF diagnosis; detection; electrograms; triggers include diagnosis; therapy; atrial episode; PMT termination; PC shock delivery; noise reversion; magnet reversion; morphology template verification; lead noise detected, non-sustained lead noise detected, NSVT/NSVF

Histogram: Peak Filtered Rate Histogram: Atrial Heart Rate Histogram:

Therapy Summary Diagram of therapies delivered

Episodes Summary Directory listing of up to 60 episodes with access to more details including stored electrograms

Lifetime Diagnostics History of bradycardia events and device-initiated charging

AT/AF Burden Trend Ventricular HV Lead Impedance Trend Trend data and counts

Multi-Vector Trend Data

Event Histogram; AV Interval Histogram; Mode Switch Duration

Ventricular Heart Rate Histogram; AT/AF Burden; Exercise and Activity

Trending; V Rates During AMS, DirectTrend™ reports up to 1 year Information regarding PMT detections Pacing lead impedances; high-voltage lead impedances; and signal Real-Time Measurements (RTM)

amplitudes On: Off

CorVue™ Congestion Monitoring CorVue Congestion Trigger 8-18 days

\* LV first with 10 ms interventricular delay



Customer Support: 46-8-474-4756

Brief Summary: Prior to using these devices, please review the Instructions for Use for a complete listing of indications, contraindications, warnings, precautions, potential adverse events and directions for use Devices depicted may not be available in all countries. Check with your St. Jude Medical representative for product availability in your country.

Unless otherwise noted, ™ indicates that the name is a trademark of, or licensed to, St. Jude Medical or one of its subsidiaries. ST. JUDE MEDICAL and the nine-squares symbol are registered and unregistered trademarks and service marks of St. Jude Medical, Inc. and its related companies. ©2013 St. Jude Medical, Inc. All rights reserved. Item GMCRM1074EN

